



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

### ***Devices Inc.***

**5480 La Sierra Dr. Dallas, TX 75231**

*(Hereinafter called the Organization) and hereby declares that Organization is accredited  
in accordance with the recognized International Standard:*

### **ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the  
operation of a laboratory quality management system  
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Electrical, Mechanical, Thermodynamic, and Time & Frequency Calibration***  
***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

*Initial Accreditation Date:*

April 03, 2018

*Issue Date:*

August 20, 2024

*Expiration Date:*

August 31, 2026

*Revision Date:*

April 1, 2025

*Accreditation No.:*

94948

*Certificate No.:*

L24-635-R1

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a  
continuous accreditation cycle. The validity of this certificate should be  
confirmed through the PJLA website: [www.pjlabs.com](http://www.pjlabs.com)*



# Certificate of Accreditation: Supplement

## Devices Inc.

5480 La Sierra Dr., Dallas, TX 75231  
Contact Name: Chas Gile Phone: 214-707-9075

*Accreditation is granted to the facility to perform the following calibration:*

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Simulation <sup>OF</sup> Indication and Control	-200 °C to 1 370 °C	0.34 °C	Fluke 724	PRO #1002
DC Current Output <sup>OF</sup>	4 mA to 20 mA	66 $\mu$ A	Fluke 789	PRO #1005
DC Current Input <sup>OF</sup>	4 mA to 20 mA	1.5 $\mu$ A	Fluke 789	PRO #1004

### Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Pressure Devices <sup>OF</sup>	Up to 2 psi	0.03 % of reading	Additel 681	PRO #1003
	5 to 500 psi	0.3 psi		
Inline, Insertion, or Clamp on flow meters <sup>OF</sup>	Up to 500 gpm	0.11 % of reading	Flexim Fluxus F601	PRO #1001
Volume- Open Mark Flasks <sup>OF</sup>	Up to 1000 mL	0.06 mL	Corning 3062-1L	PRO #1008

### Thermodynamics

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Measure <sup>OF</sup> Incubators, Stoves, Ovens, Environmental Chambers, refrigerators, Freezers & Similar Equipment	-30 °C to 1 000 °C	1.2 °C	Omega HH376/ T Probe	PRO #1006
Relative Humidity <sup>OF</sup>	Up to 95 % RH	2 % of Reading	Vaisala RH	PRO #1009

### Time & Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Stopwatches/timers <sup>OF</sup>	Up to 24 hrs	0.6 sec/day	Stopwatch NIST 960-12	PRO #1007



## *Certificate of Accreditation: Supplement*

### **Devices Inc.**

5480 La Sierra Dr., Dallas, TX 75231  
Contact Name: Chas Gile Phone: 214-707-9075

*Accreditation is granted to the facility to perform the following calibration:*

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location